

Rabbit Specific FITC/SABC Detection Kit – KOA0112

Rabbit Specific FITC/SABC Detection Kit is a streptavidin-biotin immuno-enzymatic antigen detection system. This system involves the sequential incubation of the sample with an unconjugated primary antibody specific to the target antigen, a biotinylated secondary antibody that reacts with the primary antibody, enzyme-labeled streptavidin, and chromogen substrate.

I. KIT COMPONENTS

Rabbit Specific FITC/DAB (ABC) Detection Kit			
Component	Catalog Number	Description	Size
Blocking Reagent	KOB0101	Normal Goat Serum Blocking Reagent (10X)	10ml
Biotinylated Secondary Antibody	KOK0603	Anti-Rabbit IgG (Goat) Biotin Conjugated Secondary Antibody	1ml (2mg/ml)
SABC-FITC	KOS0201	Fluorescein Conjugated Streptavidin (100-200X)	1ml (2mg/ml)
Dropper Bottles	KOJ0101	Dropper Bottles (empty for dilution)	3

II. STORAGE

Store at 4°C for one year. Avoid freezing.

III. INTRODUCTION

StreptAvidin-Biotin Complex (SABC) is specially designed for displaying the distribution of antigens in tissues and cells in immunochemistry and other immunodetection analyses. This kit has high sensitivity because each complex it generates has a large number of FITC and streptavidin molecules. Compared to the traditional immunodetection using free FITC dyes, the SABC-FITC system greatly enhances the sensitivity and suppresses the background. FITC excitation is at 490 - 495nm, emission at 520 - 530nm.

IV. REQUIRED EQUIPMENT AND REAGENTS

- APES or POLY-L-LYSINE
- 0.02M PBS (pH 7.2 - 7.6)
 - 8.5g sodium chloride, 2.8g anhydrous Na₂HPO₄ and 0.4g anhydrous NaH₂PO₄ in 1000ml of distilled water. (The weight should be adjusted accordingly if hydrous phosphates are used.)
- 0.01 M Citrate Buffer
 - 3.0g sodium citrate dihydrate (C₆H₅Na₃O₇·2H₂O) and 0.4g citric acid monohydrate (C₆H₅Na₃O₇·H₂O) in 1000ml of distilled water.

- 0.1% trypsinase or the compound digest solution

V. IHC PROTOCOL

Note: Rabbit IgG refers to the animal origin of the primary antibody, not the origin of the specimen. This Kit must be used on primary antibodies from rabbit.

A. Options For Immunohistochemistry Staining Process

The best process among the following may have to be identified by the end user. The characteristics of the antigen/antibody used may be followed as a guideline.

- **Heat-induced antigen retrieval process**
Heat-induced antigen retrieval applies to immunohistochemical analysis of paraffin-embedded sections to expose the antibody binding site on the antigens.
- **Enzyme digestion process**
Enzyme digestion process applies to immunohistochemical analysis of paraffin-embedded sections to expose the antibody binding site on the antigens.
- **Non-digestion/non-retrieval process**
Non-digestion/non-repair process applies to stable antigens using immunohistochemical analysis of paraffin-embedded sections.
- **Blood smear, cultured cells and frozen section staining process**
Blood smear, cultured cells and frozen section staining process applies to immunocytochemistry of blood smear and cultured cells, and immunohistochemical analysis of frozen-embedded sections.

B. Assay Procedure

Heat-induced antigen retrieval process

1. Cover the entire surface of a clean microslide with APES or POLY-L-LYSINE. Incubate for 1 minute and then rinse the microslide with water. Mount a tissue section (~5µm thick) onto the treated microslide and bake in an oven at 58-60°C for 30-60 minutes to ensure strong adhesion of the tissue section.
2. Dewax the tissue section in dimethylbenzene for 10 minutes and rinse with water.
3. For heat-induced antigen retrieval, add a sufficient volume of 0.01M citrate buffer (pH 6.0) in a microwavable container and preheat the buffer to 90°-100°C in a microwave. Place the slides into the container with the preheated citrate buffer (the buffer should cover the slides by at least a few centimeters) and place the container inside the microwave. Set the microwave to full power and boil for 15-20 minutes. Remove the container from the microwave and allow the slides to cool at room temperature for 15-20 minutes.
4. Wash the slides two times for 2 minutes each with 0.01 M PBS (pH 7.2-7.4) with gentle agitation.
5. Dilute normal goat serum blocking reagent at 1:10 with 0.02M PBS (pH7.2-7.6). Add reagent to the tissue section and incubate at room temperature for 20 minutes. Discard the blocking reagent solution, but do not wash the tissue section.
6. Add appropriately diluted primary antibody (rabbit IgG) and incubate at 37°C for 1 hour or at 20°C for 2 hours or at 4°C overnight.

7. Wash with 0.01M PBS (pH 7.2-7.4) 3 times for 2 minutes each.
 - The primary antibody concentration, incubation time and temperature directly affect the staining efficiency and background intensity. If the positive staining is too weak, the concentration of the primary antibody and the incubation time can be increased; if the background is too high, the primary antibody concentration and the incubation time can be decreased.
8. Add biotinylated goat anti-rabbit IgG and incubate at 20-37°C for 20 minutes. Wash the slides with 0.01M PBS (pH 7.2-7.4) 3 times for 2 minutes each.
9. Dilute SABC-FITC (Streptavidin-FITC) at 1:100 with 0.02M PBS (pH 7.2~7.6). Add SABC-FITC (Streptavidin-FITC) solution and incubate at 20-37°C for 20 minutes. Wash the slides 4 times with 0.01M PBS (pH 7.2-7.4) for 5 minutes each.
10. Put several drops of the water soluble sealing reagent onto the tissue section and seal with a cover slide. The tissue section is ready for observation under a fluorescence microscope.

Blood smear, cultured cells or frozen sections staining process

1. Treat a microslide with POLY-L-LYSINE as described above.
 - Blood samples: Add anticoagulant to the samples and smear the blood samples onto the treated slide.
 - Cultured cells: Cultured cells can be smeared onto or directly cultivated on the treated slide.
 - Frozen tissue sections: Sections of frozen tissue may be placed onto the treated slide and air-dry at room temperature for 30 minutes until no liquid is visible.
2. Fix the sample with 4% paraformaldehyde or 10% formalin for 60-90 minutes.
3. Dilute 30% H₂O₂ at 1:50 with pure methanol. Incubate the fixed sample for 30 minutes in the diluted H₂O₂ to quench the endogenous peroxidase activity. Wash the sample with distilled water 3 times for 2 minutes each.

If the direct staining result of frozen sections is not satisfactory, the tissue sections may be repaired by following the step 4 under antigen retrieval process.

4. Follow steps 5-11 in the heat-induced antigen retrieval process.

Note:

- If the staining background is too high, wash the section with 0.01-0.02% TWEEN 20-PBS (pH7.2-7.6) 4 times and then with PBS twice after SABC reaction and before DAB staining, followed by DAB chromogenic Kit to stain the section.
- 0.01M citrate buffer (pH 6.0), PBS, or PBS buffer may be used to repair the section.

VI. TROUBLESHOOTING

Weak or No Signal

Possible Cause	Solution
Slides lose signal over time during storage	Prepare slides with freshly-sectioned tissues Store slides at 4°C Do not bake slides before storage
The antibody used is not suitable for IHC procedures which detect proteins in its native conformation	Check the antibody datasheet to make certain that it has been validated for IHC applications Check if the antibody is applicable for the right IHC

	<p>samples (paraffin sections vs. frozen samples) Perform Western blot in both its native and denatured forms to ensure that the antibody detects the native form</p>
Fixation procedures (using formalin/paraformaldehyde fixatives) have masked the epitope that the antibody recognizes	<p>Use different antigen retrieval methods to unmask the epitope (HIER or PIER) Fix the sections in a shorter time</p>
The primary and/or secondary antibody has lost its activity due to improper storage, dilution or excessive freezing and thawing	<p>Run positive controls to ensure that the primary and/or secondary antibody is working properly Store the antibodies per manufacturer instructions Avoid contamination of antibodies and exposure to light</p>
Insufficient deparaffinization	<p>Increase the deparaffinization time Use fresh dimethylbenzene</p>
The protein is located in the nucleus and the antibody cannot penetrate the nucleus	<p>Add a permeabilizing agent (e.g. Triton X) to the blocking buffer and antibody dilution buffer</p>
The PBS buffer is contaminated with bacteria that damage the phosphate groups on the protein of interest	<p>Add 0.01% azide in the PBS antibody storage buffer Use fresh sterile PBS</p>
The primary and the secondary antibodies are not compatible	<p>Use a secondary antibody that was raised against the species in which the primary was raised (e.g. if the primary antibody was raised in mouse, an anti-mouse secondary antibody should be used) Check that the isotypes of the primary and secondary antibodies are compatible</p>
The protein is not present in the tissue of interest or is not sufficiently expressed	<p>Run positive controls to ensure that target protein is present in the tissue Include an amplification step in your protocol Use higher antibody concentration</p>
Insufficient antibody to detect protein of interest	<p>Use a higher antibody concentration Incubate for a longer time (e.g. overnight at 4°C)</p>
Tissue has dried out	<p>Cover the tissue sections in liquid at all time during the experiment</p>

High Background

Possible Cause	Solution
The blocking buffer is incorrect	<p>Make sure to use the blocking buffer recommended by the manufacturer</p>
Blocking is insufficient (Do not over-block the tissue because antigenic sites may be masked)	<p>Increase blocking time Change blocking reagent: (a) For tissue sections, use 10% normal serum (1 hour) (b) For cell cultures, use 1-5% BSA (0.5 hours)</p>
The primary antibody concentration is too high	<p>Titrate the antibody to determine the optimal concentration Incubate at 4°C</p>
Non-specific binding by secondary antibody	<p>Run a secondary control without primary antibody: If you see staining with your secondary only: (a) Change secondary antibody or (b) Use secondary antibody that has been pre-adsorbed Block sample with serum from the same species as the host in which the secondary antibody was raised</p>

Endogenous peroxide or phosphatase is active	Quench the endogenous peroxidase or phosphatase activity by enzyme inhibitors: (a) Peroxidase: use H ₂ O ₂ and methanol (v/v: 0.3%: 99.7%) (b) Phosphatase: 2mM Levamisol
Tissue section is too thick for reagent penetration	Prepare thinner section
Too much substrate was applied (enzymatic detection)	Dilute substrate Reduce substrate incubation time Choose substrate of higher S/N ratio e.g. Metal-enhanced DAB
Incubation temperature is too high	Incubate samples at 4°C
Primary antibody was raised in the same species as source of tissue (therefore, secondary antibody recognizes and binds non-specifically to the tissue)	Use primary antibody raised against a species which is different from the source of tissue Use biotinylated primary antibody and conjugated streptavidin for the detection system
Secondary antibody binds endogenous IgG	Include control slide stained without the primary antibody to confirm whether the secondary antibody is the source of the background
Fixation reagents are still present (Due to insufficient tissue washing)	Wash the tissues extensively with PBS buffer
Reaction between chromogens and PBS buffer in tissue or cell samples	Before incubating with the substrate, use Tris buffer to wash the samples
Membrane damage by permeabilization	Use a less stringent detergent such as Tween 20 (instead of Triton X) Remove permeabilizing agent from your buffers
Insufficient deparaffinization	Increase the deparaffinization time Use fresh dimethylbenzene
High levels of endogenous biotin in biotin-based detection systems for samples (e.g. liver and kidney tissues)	Perform biotin block after normal blocking procedure (before primary antibody incubation) Use polymer-based detection

VII. RELATED PRODUCTS

Component	Catalog #	Size
10x PBS pH 7.5	MB-012	1000ml
10x TPBS pH 7.5	MB-013	1000ml
10X PBS pH 7.2	MB-008	1000ml
10x PBST pH 7.2	MB-075-1000	1000ml